
CERTIFICATE OF ANALYSIS

Product Name	Gymnin, Cat.#: 312301
Lot No.	MTFA32438-0214
Sequence	KTCENLADDY
Dissolution condition	100%H ₂ O
Length	10AA
Modification	N/A
Molecular Weight (MW)	1171.23
Storage	-20°C

Test Items	Specifications	Results
MW by MS	1171.30	Conforms
Purity by HPLC	>95%	96.799%
Peptide content	N/A	N/A
Moisture content	N/A	N/A
Acetic acid content	N/A	N/A
Appearance	White to off-white lyophilized powder	Conforms
Quantity	20mg	5.0mg*4

NovoPro Bioscience Inc. (hereafter NovoPro) warrants material of said quality at the time of sale. It is the sole responsibility of the customers to determine the adequacy of all materials for any intended or specific purpose or use. NovoPro's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way in or which has been misused or damaged by accident or negligence. All claims must be received writing (by fax or email) within 30 days from date when product arrive at the destination city and failure to do so shall constitute a waiver by customers for any and all such claims.

Certified by:



Quality Assurance Department

Feb/27/2025




Sample Information

Order ID :MTFA32438
 Name :Gymnin
 Sequence :KTCENLADDY
 Lot. No :MTFA32438-0214
 Pump A :0.1%Trifluoroacetic in 100% water
 Pump B :0.1%Trifluoroacetic in 100% acetonitrile
 Total Flow :1ml/min
 Wavelength :214nm
 Analytical column type :Kromasil 100-5-C18 (4.6*250mm*5 μm)
 Dissolution method :100%H2O
 Inj. Volume : 20ul

Time	Module	Action	Value
0.01	Pumps	B.Conc	15
30.00	Pumps	B.Conc	45
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	15
50.00	Controller	Stop	

Chromatogram




Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	9.246	65441	7987	1.398
2	9.523	4531417	553094	96.799
3	9.746	61653	7025	1.317
4	12.459	22735	2744	0.486
Total		4681247	570849	100.000

MS Spectrum



Sample Information

Dissolution method	:15% ACN+85% H ₂ O	Interface	:ESI	Prerod Bias	:+4.5kv
Date Acquired	:2025/02/26 10:47:46	Nebulizing Gas Flow	:1.50L/min	Detector	:-0.2kv
Injection Volume	:0.2ul	CDL Temp	:250°C	T.Flow	:0.2ml/min
Block Temp	:200	CDL Volt	:0v	B.conc	:50% H ₂ O/50% MEOH

Order ID	:MTFA32438
Name	:Gymnin
Sequence	:KTCENLADDY
Lot.No	:MTFA32438-0214
Theoretical	:1171.23
Observed	:1171.30