

---

**CERTIFICATE OF ANALYSIS**

<b>Product Name</b>	[Asn670,Leu671]-Amyloid-beta Precursor770 (667-675) peptide, Cat.#314298
<b>Lot No.</b>	GTA34376-0713
<b>Sequence</b>	SEVNLDAEF
<b>Dissolution condition</b>	15%ACN+85%H2O
<b>Length</b>	9AA
<b>Modification</b>	N/A
<b>Molecular Weight (MW)</b>	1023.04
<b>Storage</b>	-20°C

---

<b>Test Items</b>	<b>Specifications</b>	<b>Results</b>
<b>MW by MS</b>	1022.55	Conforms
<b>Purity by HPLC</b>	>95%	97.699%
<b>Peptide Content</b>	N/A	N/A
<b>Moisture content</b>	N/A	N/A
<b>Acetic acid content</b>	N/A	N/A
<b>Appearance</b>	White to off-white lyophilized powder	Conforms
<b>Quantity</b>	20mg	5.0mg*4

---

NovoPro Bioscience Inc. (hereafter NovoPro) warrants material of said quality at the time of sale. It is the sole responsibility of the customers to determine the adequacy of all materials for any intended or specific purpose or use. NovoPro's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way in or which has been misused or damaged by accident or negligence. All claims must be received writing (by fax or email) within 30 days from date when product arrive at the destination city and failure to do so shall constitute a waiver by customers for any and all such claims.

Certified by:



Quality Assurance Department

July/15/2022




Sample Information

Order ID :GTA34376  
 Name :Precursor770 (667-675) peptide,Cat.# 314298  
 Sequence :SEVNLDAEF  
 Lot. No :GTA34376-0713  
 Pump A :0.1% Trifluoroacetic in 100% water  
 Pump B :0.1% Trifluoroacetic in 100% acetonitrile  
 Total Flow :1ml/min  
 Wavelength :214nm  
 Analytical column type :YMC-Triart C18(4.6\*250mm\*5um)  
 Dissolution method :15% ACN+85% H2O  
 Inj. Volume : 35ul

Time	Module	Action	Value
0.01	Pumps	B.Conc	20
25.00	Pumps	B.Conc	45
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	20
50.00	Controller	Stop	


Chromatogram



Peak Table

Peak#	Ret. Time	Area	Height	Area%
1	11.358	1343	222	0.026
2	11.996	4288	347	0.084
3	12.224	4967173	631839	97.699
4	12.463	42818	4919	0.842
5	13.162	11383	1651	0.224
6	13.991	57165	2424	1.124
Total		5084169	641402	100.000

## MS Spectrum



## Sample Information

Dissolution method	:15% ACN+85% H <sub>2</sub> O	Interface	:ESI	Prerod Bias	:+4.5kv
Date Acquired	:2022/07/14 13:20:53	Nebulizing Gas Flow	:1.50L/min	Detector	:-0.2kv
Injection Volume	:0.2ul	CDL Temp	:250C	T.Flow	:0.2ml/min
Block Temp	:200	CDL Volt	:0v	B.conc	:50% H <sub>2</sub> O/50% MEOH

Order ID	:GTA34376
Name	:Precursor770(667-675) peptide,Cat.#314298
Sequence	:SEVNLD AEF
Lot.No	:GTA34376-0713
Theoretical	:1023.04
Observed	:1022.55